



UNIVERSIDADE FEDERAL DO RIO DE JANEIRO  
Concurso Público para provimento de vagas de cargos Técnico-Administrativos – Edital 55/2015  
Chave de Correção Preliminar da Prova Discursiva

Opção de Vaga:  
301 - Pesquisador B-C - Cromatografia e Espectrometria de Massas

QUESTÃO 1:

A)

Chave de Correção	
Resposta	Qtde de Pontos
CG×CG – Conventional GC, with two columns connected. First dimension (1D) is a conventional column (e.g. 30 m x 0.25 mm x 0.25 µm). Second dimension (2D) is a short one with reduced internal diameter, typically used in fast GC (e.g. 1-2 m x 0.1 mm x 0.1 µm). Modulator between two columns.	10
<b>Total</b>	<b>10</b>

B)

Chave de Correção	
Resposta	Qtde de Pontos
Advantages: Improvement of the peak capacity. All sample injected is analysed by both columns. Allow the graphic presentation of the chromatograms. Three dimensional representations of the results. Improvement of the sensitivity.	10
<b>Total</b>	<b>10</b>

QUESTÃO 2:

Chave de Correção	
Resposta	Qtde de Pontos
The size of the particles has direct influence in the efficiency of separation. Reduction of the Height Equivalent to a Theoretical Plate (H). High pressure. Ultra- High Performance Liquid Chromatography. Core shell or porous shell – allow a reduce in the particle diameter, controlling the increase of the pressure of the system.	20
<b>Total</b>	<b>20</b>

QUESTÃO 3:

Chave de Correção	
Resposta	Qtde de Pontos
<i>Peak skewing</i> – distorted acquisition of the mass spectrum due to a fast change of the concentration in the quadrupole ion source. Main instrumental parameters: <i>Mass range</i> : mass interval which the quadrupole is able to analyse. Reducing the mass range increase the acquisition frequency and decrease the peak skewing effect. <i>Acquisition speed</i> : define amount of spectra acquired per second. The increase of the acquisition speed reduces the peak skewing effect. <i>Dwell time</i> : time window that each ion has to cross the analyser and reach the detector. <i>Interscan time</i> : time necessary to the instrument electronics re-establish the setting before a new scan.	20
<b>Total</b>	<b>20</b>



**UNIVERSIDADE FEDERAL DO RIO DE JANEIRO**  
**Concurso Público para provimento de vagas de cargos Técnico-Administrativos – Edital 55/2015**  
**Chave de Correção Preliminar da Prova Discursiva**

**QUESTÃO 4:**

<b>Chave de Correção</b>	
<b>Resposta</b>	<b>Qtde de Pontos</b>
A) It could be considered: Electrospray ionization, Chemical ionization, electron ionization, atmospheric pressure chemical ionization, and matrix assisted laser desorption ionization. Others options could be considered if corrected. B) Q1 and q2 are time independent and space dependent; Q3is space independent andtime dependent. C) Argon, Helium and Nitrogen. Others options could be considered if corrected.	20
<b>Total</b>	<b>20</b>

**QUESTÃO 5:**

<b>Chave de Correção</b>	
<b>Resposta</b>	<b>Qtde de Pontos</b>
It will be considered correct the options: Preparation of statistical calculations, graphics, macros, reports, corporations based on statistical methods. Other applications can be accepted if correct.	20
<b>Total</b>	<b>20</b>